invitrogen USER GUIDE

# eBioscience<sup>™</sup> Annexin V-FITC Apoptosis Detection Kit

Catalog Numbers BMS500FI-20, BMS500FI-100, and BMS500FI-300

Pub. No. MAN0016665 Rev. A.0 (30)



**WARNING!** Read the Safety Data Sheets (SDSs) and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves. Safety Data Sheets (SDSs) are available from **thermofisher.com/support**.

### **Product description**

The Annexin V-FITC Kit can be used to detect phosphatidylserine on the outer leaflet of the cell membrane using flow cytometry.

### Summary

Annexins are a family of calcium-dependent phospholipid-binding proteins. They are abundant in eukaryotic organisms belonging to a family of ubiquitous cytoplasmic proteins involved in signal transduction. All annexins have been shown to have a putative binding site for protein kinase C (PKC) but only annexin V would possess a potential pseudo-substrate site. Thus annexin V seems to modulate the activity of some PKCs on their substrates.

Annexin V was found to play a major role in matrix vesicle-initiated cartilage calcification as a collage-regulated calcium channel. Annexin V binds to procoagulant phospholipids (Vascular anticoagulant  $\alpha$ ) with high affinity.

Annexin V's preferential binding partner is phosphatidylserine (PS). PS is predominantly located in membrane leaflets, which face the cytosol. However, recent findings show that each cell type has the molecular machinery to expose PS at its cell surface. This machinery is activated during the execution of apoptosis. Once PS is exposed at the cell surface it exhibits procoagulant and proinflammatory activities. Annexin V will bind to the PS-exposing apoptotic cell and can inhibit the procoagulant and proinflammatory activities of the dying cell.

# Principles of the test

Annexin V exhibits anti-phospholipase activity and binds to phosphatidylserine. FITC labelling allows simple direct detection by FACS analysis. Counterstaining by propidium iodide allows the discrimination of apoptotic cells.

#### Reagents provided

Reagents	BMS500FI-20	BMS500FI-100	BMS500FI-300
	(for 20 tests)	(for 100 tests)	(for 300 tests)
rh Annexin V-	1 vial	1 vial	1 vial
FITC	(100 μL)	(500 μL)	(1.5 mL)
Binding Buffer (4x)	1 bottle	1 bottle	1 bottle
	(50 mL)	(50 mL)	(50 mL)
Propidium Iodide	1 vial	1 vial	2 vials
(20 µg/mL)	(1.8 mL)	(1.8 mL)	(1.8 mL)

## Storage and stability

Store kit reagents between 2° and 8°C. Immediately after use remaining reagents should be returned to cold storage (2–8°C).

Expiry of the kit and reagents is stated on labels. Expiry of the kit components can be guaranteed only if the components are stored properly, and if, in case of repeated use of one component, this reagent is not contaminated by the first handling.

## Materials required but not provided

- 5 mL and 10 mL graduated pipettes
- 5  $\mu L$  to 1,000  $\mu L$  adjustable single channel micropipettes with disposable tips
- · Beakers, flasks, cylinders necessary for preparation of reagents
- Glass-distilled or deionized water
- Bench top centrifuge
- Flow Cytometer
- PBS (for 1 liter: 8.00 g NaCl, 0.20 g KCl, 1.78 g Na<sub>2</sub>HPO<sub>4</sub> x 2 H<sub>2</sub>O, 0.27 g KH<sub>2</sub>PO<sub>4</sub>)

#### Precautions for use

- All reagents should be considered as potentially hazardous. We
  therefore recommend that this product is handled only by those
  persons who have been trained in laboratory techniques and that it
  is used in accordance with the principles of good laboratory
  practice. Wear suitable protective clothing such as laboratory
  overalls, safety glasses and gloves. Care should be taken to avoid
  contact with skin or eyes. In the case of contact with skin or eyes
  wash immediately with water. See material safety data sheet(s)
  and/or safety statement(s) for specific advice.
- Reagents are intended for research use only and are not for use in diagnostic or therapeutic procedures.
- Do not mix or substitute reagents with those from other lots or other sources.
- Do not use kit reagents beyond expiration date on label.
- Do not expose kit reagents to strong light during storage or incubation.
- Do not pipette by mouth.
- Do not eat or smoke in areas where kit reagents or samples are handled.
- Avoid contact of skin or mucous membranes with kit reagents or samples.
- Rubber or disposable latex gloves should be worn while handling kit reagents or samples.
- Avoid splashing or generation of aerosols.
- To avoid microbial contamination or cross-contamination of reagents or samples that may invalidate the test, use disposable pipette tips and/or pipettes.
- Glass-distilled water or deionized water must be used for reagent preparation.
- Decontaminate and dispose samples and all potentially contaminated materials as if they could contain infectious agents.
   The preferred method of decontamination is autoclaving for a minimum of 1 hour at 121.5°C.
- Liquid wastes not containing acid and neutralized waste may be mixed with sodium hypochlorite in volumes such that the final mixture contains 1.0% sodium hypochlorite. Allow 30 minutes for effective decontamination. Liquid waste containing acid must be neutralized prior to the addition of sodium hypochlorite.

#### Preparation of reagents

Dilute Binding Buffer (4x) 1:4 in distilled water (50 mL binding buffer and 150 mL distilled water).



## Test protocol

- 1. Wash cells in PBS by gentle shaking or pipetting up and down.
- 2. Resuspend cells in 200  $\mu$ L Binding Buffer (1x); cell density should be 2–5x10<sup>5</sup>/mL.
- 3. Add 5 µL Annexin V-FITC to 195 µL cell suspension.
- **4.** Mix and incubate for 10 min at room temperature.
- 5. Wash cells in 200  $\mu L$  Binding Buffer (1x) and resuspend in 190  $\mu L$  Binding buffer (1x).
- 6. Add 10 μL Propidium Iodide (20 μg/mL)
- 7. Perform FACS analysis.

#### Limitations

 Bacterial or fungal contamination of either screen samples or reagents or cross-contamination between reagents may cause erroneous results.  Disposable pipette tips, flasks or glassware are preferred, reusable glassware must be washed and thoroughly rinsed of all detergent before use.

#### Limited product warranty

Life Technologies Corporation and/or its affiliate(s) warrant their products as set forth in the Life Technologies' General Terms and Conditions of Sale found on Life Technologies' website at www.thermofisher.com/us/en/home/global/terms-and-conditions.html. If you have any questions, please contact Life Technologies at www.thermofisher.com/support.



Manufacturer: Bender MedSystems GmbH | Campus Vienna Biocenter 2 | 1030 Vienna, Austria

The information in this guide is subject to change without notice.

DISCLAIMER: TO THE EXTENT ALLOWED BY LAW, LIFE TECHNOLOGIES AND/OR ITS AFFILIATE(S) WILL NOT BE LIABLE FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE, OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING YOUR USE OF IT.

Important Licensing Information: These products may be covered by one or more Limited Use Label Licenses. By use of these products, you accept the terms and conditions of all applicable Limited Use Label Licenses.

©2018 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified. All other trademarks are the property of their respective owners.

